

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/02/2024

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:08
INTERNAL STANDARD	VERIFIED	10:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12

*Donnie J Long* *844*

