

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/08/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:11
INTERNAL STANDARD	VERIFIED	09:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:14

*DOA #427*

