

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/22/2024

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:50
INTERNAL STANDARD	VERIFIED	15:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:50
BLANK TEST	0.00	15:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:52
BLANK TEST	0.00	15:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:53
BLANK TEST	0.00	15:54

*SS Allen #759*

