

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:36
INTERNAL STANDARD	VERIFIED	15:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:38
BLANK TEST	0.00	15:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:40

J Riehl
#820

