

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/15/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:11
BLANK TEST	0.00	08:12

JM Fish #427

