

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:09
INTERNAL STANDARD	VERIFIED	13:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:10
BLANK TEST	0.00	13:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:11
BLANK TEST	0.00	13:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:12
BLANK TEST	0.00	13:13

[Handwritten signature]
#820

