

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/01/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:51
INTERNAL STANDARD	VERIFIED	04:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:52
BLANK TEST	0.00	04:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:53
BLANK TEST	0.00	04:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:54
BLANK TEST	0.00	04:55

*[Handwritten signature]*  
#820

