

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/01/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:07
INTERNAL STANDARD	VERIFIED	16:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:08
BLANK TEST	0.00	16:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:10
BLANK TEST	0.00	16:10

*PC Fish # 427*

