

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/09/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:19
INTERNAL STANDARD	VERIFIED	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:20
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:21
BLANK TEST	0.00	10:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:22
BLANK TEST	0.00	10:23

J Riehl
#520

