

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/01/2024

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:43
INTERNAL STANDARD	VERIFIED	10:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:45
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:46
BLANK TEST	0.00	10:47

