

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/08/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:01
INTERNAL STANDARD	VERIFIED	08:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:02
BLANK TEST	0.00	08:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:04
BLANK TEST	0.00	08:05

*PO JMF #407*

