

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/01/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:22
INTERNAL STANDARD	VERIFIED	13:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:24
BLANK TEST	0.00	13:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:25
BLANK TEST	0.00	13:26

Handwritten signature: J. M. Fish #427

