

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:24
INTERNAL STANDARD	VERIFIED	11:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:25
BLANK TEST	0.00	11:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:26
BLANK TEST	0.00	11:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:27
BLANK TEST	0.00	11:28

J Riehl
#870

