

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:49
INTERNAL STANDARD	VERIFIED	10:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:50
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53

J Riehl
#520

