

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/29/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:19
INTERNAL STANDARD	VERIFIED	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:20
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:22
BLANK TEST	0.00	10:22

*PO Fish #427*

