

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/10/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:55
INTERNAL STANDARD	VERIFIED	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:59

