

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/08/2024

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:24
INTERNAL STANDARD	VERIFIED	14:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:24
BLANK TEST	0.00	14:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:26
BLANK TEST	0.00	14:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:27
BLANK TEST	0.00	14:28

