

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/15/2024

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:34
INTERNAL STANDARD	VERIFIED	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:35
BLANK TEST	0.00	11:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:36
BLANK TEST	0.00	11:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:37
BLANK TEST	0.00	11:38

