

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:51
INTERNAL STANDARD	VERIFIED	01:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55

820

