

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/29/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:37
INTERNAL STANDARD	VERIFIED	15:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:38
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:39
BLANK TEST	0.00	15:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:40
BLANK TEST	0.00	15:41

*POA #27*

