

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/01/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:46
INTERNAL STANDARD	VERIFIED	23:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:47
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:49
BLANK TEST	0.00	23:50

*PO Fish #427*

