

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/09/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:33

[Handwritten signature]
#820

