

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:28
INTERNAL STANDARD	VERIFIED	22:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32

J Riehl
#520

