

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208




Date: 12/16/2024

OPERATOR NAME:
BRANDON D KINDERMAN
Permit #: 38773
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:01
INTERNAL STANDARD	VERIFIED	16:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:02
BLANK TEST	0.00	16:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:03
BLANK TEST	0.00	16:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:04
BLANK TEST	0.00	16:05

P.D. Fisher #726



SET [Signature] #757