

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 12/27/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:32
BLANK TEST	0.00	10:33

*POA #427*

