

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/15/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:46
INTERNAL STANDARD	VERIFIED	11:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:47
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:49
BLANK TEST	0.00	11:50


#820

