

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/22/2024

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:13
INTERNAL STANDARD	VERIFIED	16:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:13
BLANK TEST	0.00	16:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:16
BLANK TEST	0.00	16:17

PO Steven Bunio #789

