

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/30/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:14
INTERNAL STANDARD	VERIFIED	14:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18

J Riehl
#830

