

SUPERVISOR TEST

WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606

Date: 02/08/2024

OPERATOR NAME:  
 J RIEHL  
 Permit #: 47503  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:45
INTERNAL STANDARD	VERIFIED	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49

*[Handwritten signature]*  
 #820

