

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/22/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:15
INTERNAL STANDARD	VERIFIED	09:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19

Donnie J Long 244

