

SUPERVISOR TEST


WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 03/15/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:12
INTERNAL STANDARD	VERIFIED	09:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16



#820

