

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/22/2024

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:36
INTERNAL STANDARD	VERIFIED	19:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:36
BLANK TEST	0.00	19:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:37
BLANK TEST	0.00	19:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:39
BLANK TEST	0.00	19:40

PO Steven Bunio #789

