

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/08/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:37
INTERNAL STANDARD	VERIFIED	18:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:38
BLANK TEST	0.00	18:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:39
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41

  
#870

