

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/15/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:14

PO Fish #427

