

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:13
INTERNAL STANDARD	VERIFIED	12:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:16
BLANK TEST	0.00	12:16

[Handwritten signature]
#820

