

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/29/2024

OPERATOR NAME:  
SEAN T SMITH  
Permit #: 46227  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:05
INTERNAL STANDARD	VERIFIED	09:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:08
BLANK TEST	0.00	09:09

*Signature* dmt

