

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/01/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:53
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57

#820

