

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/08/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:47
INTERNAL STANDARD	VERIFIED	06:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51

PO J.M. Fish #427

