

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/01/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:14
INTERNAL STANDARD	VERIFIED	20:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:17
BLANK TEST	0.00	20:18

PO A/F #4127

