

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/15/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10

*POA Fish #427*

