

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/28/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:08
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12

*Handwritten signature and initials*

