

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/01/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:00
INTERNAL STANDARD	VERIFIED	12:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:01
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:02
BLANK TEST	0.00	12:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:03
BLANK TEST	0.00	12:04

