

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/08/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:13
INTERNAL STANDARD	VERIFIED	07:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16

PO A/F #427

