

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 07/17/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06

*POA #427*

