

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/22/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:48
INTERNAL STANDARD	VERIFIED	17:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:48
BLANK TEST	0.00	17:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:50
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:51
BLANK TEST	0.00	17:52

PO JMF #427

