

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 07/29/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:49
INTERNAL STANDARD	VERIFIED	18:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:51
BLANK TEST	0.00	18:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:52
BLANK TEST	0.00	18:53

*[Handwritten signature]*  
#820

