

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/15/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:29
INTERNAL STANDARD	VERIFIED	08:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:31
BLANK TEST	0.00	08:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:32
BLANK TEST	0.00	08:33



Handwritten signature and number
#714