

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/22/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:43
INTERNAL STANDARD	VERIFIED	12:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:44
BLANK TEST	0.00	12:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:45
BLANK TEST	0.00	12:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:46
BLANK TEST	0.00	12:47

*[Handwritten signature]*  
#820

