

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/01/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:01
INTERNAL STANDARD	VERIFIED	03:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:03
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05

[Handwritten signature]
#870

