

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/15/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:33
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36

[Handwritten signature]
#820

